

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|-------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARNK-0074 |
| Location: | SOUTH BRUNSWICK POLICE | Calib. No.: | 00027 |
| Calibration File No.: | 00631 | Cert. Date: | 10/17/2013 |
| Certification File No.: | 00632 | Lin. Date: | 10/17/2013 |
| Linearity File No.: | 00633 | Soln. Date: | 01/09/2014 |
| Solution File No.: | 00654 | File Date: | 01/09/2014 |
| Sequential File No.: | 00654 | | |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUF S3-0060 |
| Solution Control Lot: | 13G115 | Expires: | 07/18/2015 |
| | | Bottle No.: | 0668 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 16:02S | | |
| Control 1 EC | 0.101% | 16:02S | 34.1°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 16:02S | 34.1°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:03S | | |
| Control 2 EC | 0.100% | 16:04S | 34.1°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 16:04S | 34.1°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:04S | | |
| Control 3 EC | 0.100% | 16:05S | 34.1°C | *** TEST PASSED *** |
| Control 3 IR | 0.102% | 16:05S | 34.1°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:06S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

SGT. Gary A. #351

Badge No.: 351

Date: 01/09/2014