

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00195	Calib. Date:	08/15/2008
Certification File No.:	00196	Cert. Date:	08/15/2008
Linearity File No.:	00197	Lin. Date:	08/15/2008
Solution File No.:	00198	Soln. Date:	08/15/2008
Sequential File No.:	00198	File Date:	08/15/2008
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	08A049	Expires:	01/09/2010
		Bottle No.:	1017

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:48D		
Control 1 EC	0.101%	11:49D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:49D		
Control 2 EC	0.101%	11:50D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51D		
Control 3 EC	0.102%	11:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDSFP2-0417 TJS

Changed By:

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: TPR. I Thomas J. Snyder #5792

Badge No.: 5792

Date: 08/15/2008

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00127	Calib. Date: 09/21/2007
Certification File No.:	00128	Calib. No.: 00006
Linearity File No.:	00129	Cert. Date: 09/21/2007
Solution File No.:	00186	Cert. No.: 00004
Sequential File No.:	00186	Lin. Date: 09/21/2007
		Lin. No.: 00004
		Soln. Date: 06/17/2008
		Soln. No.: 00053
		File Date: 06/17/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	07B044	Expires: 02/20/2009
		Bottle No.: 0721

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:19D		
Control 1 EC	0.101%	15:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21D		
Control 2 EC	0.099%	15:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22D		
Control 3 EC	0.099%	15:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDSFP2-0417 JGS

Changed By:

Last Name: SNYDER	First Name: THOMAS	MI: J.
Signature: <u>TR. I Thomas J. Snyder #5792</u>	Badge No.: 5792	Date: 06/17/2008

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00127 Calib. Date: 09/21/2007 Calib. No.: 00006
Certification File No.: 00128 Cert. Date: 09/21/2007 Cert. No.: 00004
Linearity File No.: 00129 Lin. Date: 09/21/2007 Lin. No.: 00004
Solution File No.: 00183 Soln. Date: 05/23/2008 Soln. No.: 00052
Sequential File No.: 00183 File Date: 05/23/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 02/20/2009
Solution Control Lot: 07B044 Bottle No.: 0711

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:44D		
Control 1 EC	0.101%	13:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46D		
Control 2 EC	0.099%	13:46D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:46D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47D		
Control 3 EC	0.100%	13:48D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:48D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 05/23/2008

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00127 Calib. Date: 09/21/2007 Calib. No.: 00006
Certification File No.: 00128 Cert. Date: 09/21/2007 Cert. No.: 00004
Linearity File No.: 00129 Lin. Date: 09/21/2007 Lin. No.: 00004
Solution File No.: 00180 Soln. Date: 04/26/2008 Soln. No.: 00051
Sequential File No.: 00180 File Date: 04/26/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 02/20/2009
Solution Control Lot: 07B044 Bottle No.: 0701

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:43D		
Control 1 EC	0.101%	23:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:44D		
Control 2 EC	0.100%	23:45D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	23:45D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:45D		
Control 3 EC	0.099%	23:46D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:46D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 04/26/2008

Calibrating Unit New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00127	Calib. Date: 09/21/2007
Certification File No.:	00128	Calib. No.: 00006
Linearity File No.:	00129	Cert. Date: 09/21/2007
Solution File No.:	00175	Cert. No.: 00004
Sequential File No.:	00175	Lin. Date: 09/21/2007
		Lin. No.: 00004
		Soln. Date: 03/31/2008
		Soln. No.: 00048
		File Date: 03/31/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	07B044	Expires: 02/20/2009
		Bottle No.: 0703

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:51S		
Control 1 EC	0.101%	11:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52S		
Control 2 EC	0.101%	11:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:54S		
Control 3 EC	0.100%	11:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS		First Name: MICHAEL	MI: J
Signature: _____		Badge No.: 258	Date: 03/31/2008

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00127 Calib. Date: 09/21/2007 Calib. No.: 00006
Certification File No.: 00128 Cert. Date: 09/21/2007 Cert. No.: 00004
Linearity File No.: 00129 Lin. Date: 09/21/2007 Lin. No.: 00004
Solution File No.: 00168 Soln. Date: 03/07/2008 Soln. No.: 00047
Sequential File No.: 00168 File Date: 03/07/2008

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 02/20/2009
Solution Control Lot: 07B044 Bottle No.: 0712

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:43S		
Control 1 EC	0.101%	16:44S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:44S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:44S		
Control 2 EC	0.100%	16:45S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:45S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:46S		
Control 3 EC	0.100%	16:46S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:46S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 03/07/2008

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00127	Calib. Date: 09/21/2007
Certification File No.:	00128	Calib. No.: 00006
Linearity File No.:	00129	Cert. Date: 09/21/2007
Solution File No.:	00160	Cert. No.: 00004
Sequential File No.:	00160	Lin. Date: 09/21/2007
		Lin. No.: 00004
		Soln. Date: 02/07/2008
		Soln. No.: 00046
		File Date: 02/07/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	07B044	Expires: 02/20/2009
		Bottle No.: 0718

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:15S		
Control 1 EC	0.101%	15:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17S		
Control 2 EC	0.099%	15:17S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:17S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18S		
Control 3 EC	0.100%	15:19S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:19S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 258

Date: 02/07/2008