

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00305	Calib. Date: 08/26/2011
Certification File No.:	00306	Calib. No.: 00021
Linearity File No.:	00307	Cert. Date: 08/26/2011
Solution File No.:	00313	Cert. No.: 00013
Sequential File No.:	00313	Lin. Date: 08/26/2011
		Lin. No.: 00013
		Soln. Date: 01/11/2012
		Soln. No.: 00098
		File Date: 01/11/2012

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %: 0.100%		Expires: 12/16/2012
Solution Control Lot: 10L084		Bottle No.: 0131

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:08S		
Control 1 EC	0.104%	14:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:09S		
Control 2 EC	0.102%	14:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	14:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:10S		
Control 3 EC	0.101%	14:11S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HQLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 01/11/2012