

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SOUTH BRUNSWICK POLICE
Serial No.: ARNK-0074
Calibration File No.: 00430
Calib. Date: 08/26/2011
Calib. No.: 00018
Certification File No.: 00431
Cert. Date: 08/26/2011
Cert. No.: 00012
Linearity File No.: 00432
Lin. Date: 08/26/2011
Lin. No.: 00012
Solution File No.: 00457
Soln. Date: 01/11/2012
Soln. No.: 00109
Sequential File No.: 00457
File Date: 01/11/2012

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 10L084
Model No.: CU-34
Serial No.: DDUF S3-0061
Expires: 12/16/2012
Bottle No.: 0126

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:01S		
Control 1 EC	0.103%	14:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02S		
Control 2 EC	0.102%	14:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	14:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:04S		
Control 3 EC	0.102%	14:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	14:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 01/11/2012