

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00433	Calib. Date:	08/20/2015
Certification File No.:	00434	Cert. No.:	00032
Linearity File No.:	00435	Cert. Date:	08/20/2015
Solution File No.:	00441	Lin. No.:	00023
Sequential File No.:	00441	Lin. Date:	08/20/2015
		Soln. No.:	00166
		Soln. Date:	01/12/2016
		File Date:	01/12/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	0077

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:39S		
Control 1 EC	0.101%	11:40S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:40S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:40S		
Control 2 EC	0.100%	11:41S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:41S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42S		
Control 3 EC	0.100%	11:42S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:42S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 01/12/2016