

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00785 Calib. Date: 08/20/2015 Calib. No.: 00031
Certification File No.: 00786 Cert. Date: 08/20/2015 Cert. No.: 00021
Linearity File No.: 00787 Lin. Date: 08/20/2015 Lin. No.: 00021
Solution File No.: 00822 Soln. Date: 01/12/2016 Soln. No.: 00174
Sequential File No.: 00822 File Date: 01/12/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 0076

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:10S		
Control 1 EC	0.101%	09:11S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:11S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:11S		
Control 2 EC	0.101%	09:12S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:12S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13S		
Control 3 EC	0.100%	09:14S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:14S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 01/12/2016