

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTL-0014 |
| Location: | SOUTH BRUNSWICK POLICE | |
| Calibration File No.: | 00346 | Calib. No.: 00024 |
| Certification File No.: | 00347 | Cert. No.: 00016 |
| Linearity File No.: | 00348 | Lin. No.: 00016 |
| Solution File No.: | 00351 | Soln. No.: 00118 |
| Sequential File No.: | 00351 | File Date: 01/15/2013 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUH S3-0079 |
| Solution Control Lot: | 12F102 | Expires: 06/28/2014 |
| | | Bottle No.: 0408 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 14:21S | | |
| Control 1 EC | 0.100% | 14:22S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 14:22S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:23S | | |
| Control 2 EC | 0.098% | 14:23S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 14:23S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:24S | | |
| Control 3 EC | 0.098% | 14:25S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 14:25S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:25S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Badge No.: 351

Date: 01/15/2013