

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00531	Calib. Date: 10/22/2012
Certification File No.:	00532	Calib. No.: 00025
Linearity File No.:	00533	Cert. Date: 10/22/2012
Solution File No.:	00550	Cert. No.: 00015
Sequential File No.:	00550	Lin. Date: 10/22/2012
		Lin. No.: 00015
		Soln. Date: 01/15/2013
		Soln. No.: 00128
		File Date: 01/15/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0407

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:16S		
Control 1 EC	0.100%	14:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:17S		
Control 2 EC	0.099%	14:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:18S		
Control 3 EC	0.099%	14:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Badge No.: 351

Date: 01/15/2013