

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00705	Calib. Date: 09/30/2014
Certification File No.:	00706	Calib. No.: 00029
Linearity File No.:	00707	Cert. Date: 09/30/2014
Solution File No.:	00735	Cert. No.: 00019
Sequential File No.:	00735	Lin. Date: 09/30/2014
		Lin. No.: 00019
		Soln. Date: 01/20/2015
		Soln. No.: 00159
		File Date: 01/20/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0040

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:50S		
Control 1 EC	0.101%	08:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51S		
Control 2 EC	0.101%	08:52S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53S		
Control 3 EC	0.100%	08:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt Gary Holsten #351

Badge No.: 351

Date: 01/20/2015