

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00012
Calibration File No.:	00307	Calib. Date:	12/21/2009
Certification File No.:	00308	Cert. Date:	12/21/2009
Linearity File No.:	00309	Lin. Date:	12/21/2009
Solution File No.:	00313	Soln. Date:	01/21/2010
Sequential File No.:	00313	File Date:	01/21/2010

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0061
Control Solution %:	0.100%	Expires:	12/01/2010	Bottle No.:	1062
Solution Control Lot:	08L061				

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:45S		
Control 1 EC	0.101%	12:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47S		
Control 2 EC	0.101%	12:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48S		
Control 3 EC	0.101%	12:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERRS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 01/21/2010