

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00014
Calibration File No.:	00376	Cert. Date:	12/03/2010
Certification File No.:	00377	Lin. Date:	12/03/2010
Linearity File No.:	00378	Soln. Date:	01/25/2011
Solution File No.:	00390	File Date:	01/25/2011
Sequential File No.:	00390		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0060
Solution Control Lot:	091069	Expires:	09/18/2011
		Bottle No.:	0385

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:41S		
Control 1 EC	0.103%	13:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	13:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42S		
Control 2 EC	0.101%	13:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:44S		
Control 3 EC	0.101%	13:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Badge No.: 351

Date: 01/25/2011