

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00210 Calib. Date: 01/06/2009 Calib. No.: 00009  
Certification File No.: 00211 Cert. Date: 01/06/2009 Cert. No.: 00007  
Linearity File No.: 00212 Lin. Date: 01/06/2009 Lin. No.: 00007  
Solution File No.: 00214 Soln. Date: 02/06/2009 Soln. No.: 00055  
Sequential File No.: 00214 File Date: 02/06/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
Control Solution %: 0.100% Expires: 01/09/2010  
Solution Control Lot: 08A049 Bottle No.: 1014

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:45S		
Control 1 EC	0.104%	17:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:46S		
Control 2 EC	0.101%	17:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:48S		
Control 3 EC	0.101%	17:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_

Badge No.: 258

Date: 02/06/2009

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00210 Calib. Date: 01/06/2009 Calib. No.: 00009  
Certification File No.: 00211 Cert. Date: 01/06/2009 Cert. No.: 00007  
Linearity File No.: 00212 Lin. Date: 01/06/2009 Lin. No.: 00007  
Solution File No.: 00216 Soln. Date: 03/04/2009 Soln. No.: 00056  
Sequential File No.: 00216 File Date: 03/04/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
Control Solution %: 0.100% Expires: 01/09/2010  
Solution Control Lot: 08A049 Bottle No.: 1009

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:24S		
Control 1 EC	0.100%	13:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:25S		
Control 2 EC	0.098%	13:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:27S		
Control 3 EC	0.099%	13:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 258

Date: 03/04/2009