

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00379	Calib. Date: 10/17/2013
Certification File No.:	00380	Calib. No.: 00026
Linearity File No.:	00381	Cert. Date: 10/17/2013
Solution File No.:	00388	Cert. No.: 00018
Sequential File No.:	00388	Lin. Date: 10/17/2013
		Lin. No.: 00018
		Soln. Date: 02/07/2014
		Soln. No.: 00135
		File Date: 02/07/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1350

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:29S		
Control 1 EC	0.102%	14:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31S		
Control 2 EC	0.099%	14:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32S		
Control 3 EC	0.100%	14:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary A. 351

Badge No.: 351

Date: 02/07/2014