

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00631	Calib. Date:	10/17/2013
Certification File No.:	00632	Cert. Date:	10/17/2013
Linearity File No.:	00633	Lin. Date:	10/17/2013
Solution File No.:	00663	Soln. Date:	02/07/2014
Sequential File No.:	00663	File Date:	02/07/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0060
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1349

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:25S		
Control 1 EC	0.101%	14:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:27S		
Control 2 EC	0.099%	14:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:28S		
Control 3 EC	0.100%	14:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary Holsten #351

Badge No.: 351

Date: 02/07/2014