

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00022
Calibration File No.:	00314	Cert. Date:	01/16/2012
Certification File No.:	00315	Lin. Date:	01/16/2012
Linearity File No.:	00316	Soln. Date:	02/09/2012
Solution File No.:	00318	File Date:	02/09/2012
Sequential File No.:	00318		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0060
Solution Control Lot:	11I095	Expires:	09/15/2013
		Bottle No.:	0440

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:45S		
Control 1 EC	0.102%	13:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 2 EC	0.101%	13:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		
Control 3 EC	0.101%	13:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Badge No.: 351

Date: 02/09/2012