

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00024
Calibration File No.:	00346	Cert. Date:	12/26/2012
Certification File No.:	00347	Lin. Date:	12/26/2012
Linearity File No.:	00348	Soln. Date:	02/12/2013
Solution File No.:	00353	File Date:	02/12/2013
Sequential File No.:	00353		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0410

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:39S		
Control 1 EC	0.100%	11:40S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.009%	11:41S		
Control 2 EC	0.095%	11:41S	33.9°C	AMBIENT AIR CHECK ERR
Control 2 IR	-.---%	11:41S	33.9°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	11:42S		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 02/12/2013

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SOUTH BRUNSWICK POLICE
Serial No.: ARTL-0014
Calibration File No.: 00346
Calib. Date: 12/26/2012
Calib. No.: 00024
Certification File No.: 00347
Cert. Date: 12/26/2012
Cert. No.: 00016
Linearity File No.: 00348
Lin. Date: 12/26/2012
Lin. No.: 00016
Solution File No.: 00354
Soln. Date: 02/12/2013
Soln. No.: 00120
Sequential File No.: 00354
File Date: 02/12/2013

Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDUH S3-0079
Control Solution %: 0.100%
Expires: 06/28/2014
Solution Control Lot: 12F102
Bottle No.: 0411

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:20S		
Control 1 EC	-.---%	13:21S	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	13:21S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 02/12/2013