

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00025
Calibration File No.:	00531	Cert. Date:	10/22/2012
Certification File No.:	00532	Lin. Date:	10/22/2012
Linearity File No.:	00533	Soln. Date:	02/12/2013
Solution File No.:	00551	File Date:	02/12/2013
Sequential File No.:	00551		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0409

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:30S		
Control 1 EC	0.100%	11:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		
Control 2 EC	0.100%	11:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32S		
Control 3 EC	0.099%	11:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_

*[Handwritten Signature]*  
#351

Badge No.: 351

Date: 02/12/2013