

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00307 Calib. Date: 12/21/2009 Calib. No.: 00012  
Certification File No.: 00308 Cert. Date: 12/21/2009 Cert. No.: 00008  
Linearity File No.: 00309 Lin. Date: 12/21/2009 Lin. No.: 00008  
Solution File No.: 00318 Soln. Date: 02/19/2010 Soln. No.: 00078  
Sequential File No.: 00318 File Date: 02/19/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 12/01/2010  
Solution Control Lot: 08L061 Bottle No.: 1061

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:05S		
Control 1 EC	0.102%	11:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:06S		
Control 2 EC	0.101%	11:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08S		
Control 3 EC	0.101%	11:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: \_\_\_\_\_

*PFC M. Rogers*

Badge No.: 258

Date: 02/19/2010