

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00272	Calib. Date: 12/03/2010
Certification File No.:	00273	Calib. No.: 00017
Linearity File No.:	00274	Cert. Date: 12/03/2010
Solution File No.:	00282	Cert. No.: 00011
Sequential File No.:	00282	Lin. Date: 12/03/2010
		Lin. No.: 00011
		Soln. Date: 02/22/2011
		Soln. No.: 00084
		File Date: 02/22/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	091069	Expires: 09/18/2011
		Bottle No.: 0380

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:55S		
Control 1 EC	0.102%	13:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.008%	13:56S		
Control 2 EC	0.092%	13:57S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.093%	13:57S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	13:58S		
Control 3 EC	0.095%	13:58S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.091%	13:58S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	13:59S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 02/22/2011