

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00376 Calib. Date: 12/03/2010 Calib. No.: 00014
Certification File No.: 00377 Cert. Date: 12/03/2010 Cert. No.: 00010
Linearity File No.: 00378 Lin. Date: 12/03/2010 Lin. No.: 00010
Solution File No.: 00395 Soln. Date: 02/22/2011 Soln. No.: 00093
Sequential File No.: 00395 File Date: 02/22/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 09/18/2011
Solution Control Lot: 091069 Bottle No.: 0379

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:49S		
Control 1 EC	0.103%	13:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	13:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.008%	13:51S		
Control 2 EC	0.095%	13:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	13:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:52S		
Control 3 EC	0.095%	13:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.095%	13:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature:  #351

Badge No.: 351

Date: 02/22/2011