

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00376	Calib. Date: 12/03/2010
Certification File No.:	00377	Cert. No.: 00014
Linearity File No.:	00378	Cert. Date: 12/03/2010
Solution File No.:	00396	Lin. No.: 00010
Sequential File No.:	00396	Lin. Date: 12/03/2010
		Soln. No.: 00094
		Soln. Date: 03/01/2011
		File Date: 03/01/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0130

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:42S		
Control 1 EC	0.104%	09:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43S		
Control 2 EC	0.101%	09:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	09:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45S		
Control 3 EC	0.100%	09:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: \_\_\_\_\_

T.P.N. I [Signature] #5925

Badge No.: 5925

Date: 03/01/2011