

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00314	Calib. Date: 01/16/2012
Certification File No.:	00315	Calib. No.: 00022
Linearity File No.:	00316	Cert. Date: 01/16/2012
Solution File No.:	00319	Cert. No.: 00014
Sequential File No.:	00319	Lin. Date: 01/16/2012
		Lin. No.: 00014
		Soln. Date: 03/07/2012
		Soln. No.: 00101
		File Date: 03/07/2012

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUF S3-0060
Control Solution %: 0.100%		Expires: 09/15/2013
Solution Control Lot: 11I095		Bottle No.: 0449

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:28S		
Control 1 EC	0.102%	09:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29S		
Control 2 EC	0.101%	09:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31S		
Control 3 EC	0.101%	09:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_



Badge No.: 351

Date: 03/07/2012