

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00459 Calib. Date: 01/16/2012 Calib. No.: 00019
Certification File No.: 00460 Cert. Date: 01/16/2012 Cert. No.: 00013
Linearity File No.: 00461 Lin. Date: 01/16/2012 Lin. No.: 00013
Solution File No.: 00469 Soln. Date: 03/07/2012 Soln. No.: 00112
Sequential File No.: 00469 File Date: 03/07/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 09/15/2013
Solution Control Lot: 11I095 Bottle No.: 0430

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:21S		
Control 1 EC	0.101%	09:22S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:23S		
Control 2 EC	0.100%	09:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:24S		
Control 3 EC	0.100%	09:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature:  #351

Badge No.: 351

Date: 03/07/2012