

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00379	Calib. Date:	10/17/2013
Certification File No.:	00380	Cert. Date:	10/17/2013
Linearity File No.:	00381	Lin. Date:	10/17/2013
Solution File No.:	00389	Soln. Date:	03/11/2014
Sequential File No.:	00389	File Date:	03/11/2014

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0080
Control Solution %:	0.100%			Expires:	11/12/2015
Solution Control Lot:	13K125			Bottle No.:	1348

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:43S		
Control 1 EC	0.101%	15:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45S		
Control 2 EC	0.099%	15:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46S		
Control 3 EC	0.099%	15:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 03/11/2014