

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00631 Calib. Date: 10/17/2013 Calib. No.: 00027
Certification File No.: 00632 Cert. Date: 10/17/2013 Cert. No.: 00017
Linearity File No.: 00633 Lin. Date: 10/17/2013 Lin. No.: 00017
Solution File No.: 00668 Soln. Date: 03/11/2014 Soln. No.: 00146
Sequential File No.: 00668 File Date: 03/11/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 11/12/2015
Solution Control Lot: 13K125 Bottle No.: 1347

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:40S		
Control 1 EC	0.101%	15:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:41S		
Control 2 EC	0.101%	15:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:43S		
Control 3 EC	0.100%	15:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary A. #351

Badge No.: 351

Date: 03/11/2014