

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C		Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE		
Calibration File No.: 00443	Calib. Date: 02/17/2016	Calib. No.: 00033
Certification File No.: 00444	Cert. Date: 02/17/2016	Cert. No.: 00024
Linearity File No.: 00445	Lin. Date: 02/17/2016	Lin. No.: 00024
Solution File No.: 00450	Soln. Date: 03/16/2016	Soln. No.: 00170
Sequential File No.: 00450	File Date: 03/16/2016	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %: 0.100%		Expires: 08/04/2017
Solution Control Lot: 15H140		Bottle No.: 0083

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:22S		
Control 1 EC	0.099%	12:23S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:23S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23S		
Control 2 EC	0.098%	12:24S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:24S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25S		
Control 3 EC	0.097%	12:25S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:25S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 03/16/2016

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
 Location: SOUTH BRUNSWICK POLICE
 Calibration File No.: 00443
 Certification File No.: 00444
 Linearity File No.: 00445
 Solution File No.: 00449
 Sequential File No.: 00449

Serial No.: ARTL-0014

Calib. Date: 02/17/2016
 Cert. Date: 02/17/2016
 Lin. Date: 02/17/2016
 Soln. Date: 03/16/2016
 File Date: 03/16/2016

Calib. No.: 00033
 Cert. No.: 00024
 Lin. No.: 00024
 Soln. No.: 00169

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 15H140

Model No.: CU-34

Serial No.: DDUH S3-0079
 Expires: 08/04/2017
 Bottle No.: 0083

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:38S		
Control 1 EC	-.---%	09:38S	23.2°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	09:38S	23.2°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 03/16/2016