

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00244	Calib. Date: 12/21/2009
Certification File No.:	00245	Calib. No.: 00015
Linearity File No.:	00246	Cert. Date: 12/21/2009
Solution File No.:	00251	Cert. No.: 00009
Sequential File No.:	00251	Lin. Date: 12/21/2009
		Lin. No.: 00009
		Soln. Date: 03/17/2010
		Soln. No.: 00071
		File Date: 03/17/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	08L061	Expires: 12/01/2010
		Bottle No.: 1072

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:00S		
Control 1 EC	0.102%	19:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	19:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:02S		
Control 2 EC	0.100%	19:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:03S		
Control 3 EC	0.101%	19:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 03/17/2010