

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00406	Calib. Date: 09/30/2014
Certification File No.:	00407	Calib. No.: 00029
Linearity File No.:	00408	Cert. Date: 09/30/2014
Solution File No.:	00419	Cert. No.: 00021
Sequential File No.:	00419	Lin. Date: 09/30/2014
		Lin. No.: 00021
		Soln. Date: 03/17/2015
		Soln. No.: 00152
		File Date: 03/17/2015

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0034

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:02S		
Control 1 EC	0.101%	13:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	13:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04S		
Control 2 EC	0.096%	13:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:06S		
Control 3 EC	0.092%	13:06S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.101%	13:06S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	13:08S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 03/17/2015

# New Standard Solution Report

## Equipment

Alcotest 7110 MKIII-C  
Location: SOUTH BRUNSWICK POLICE  
Serial No.: ARTL-0014  
Calibration File No.: 00406 Calib. Date: 09/30/2014 Calib. No.: 00029  
Certification File No.: 00407 Cert. Date: 09/30/2014 Cert. No.: 00021  
Linearity File No.: 00408 Lin. Date: 09/30/2014 Lin. No.: 00021  
Solution File No.: 00420 Soln. Date: 03/17/2015 Soln. No.: 00153  
Sequential File No.: 00420 File Date: 03/17/2015  
Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
Control Solution %: 0.100% Expires: 08/22/2016  
Solution Control Lot: 14H132 Bottle No.: 0034

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:16S		
Control 1 EC	0.103%	14:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:17S		
Control 2 EC	0.094%	14:18S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.101%	14:18S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:19S		
Control 3 EC	0.092%	14:20S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.101%	14:20S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:21S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 03/17/2015

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00406 Calib. Date: 09/30/2014 Calib. No.: 00029  
Certification File No.: 00407 Cert. Date: 09/30/2014 Cert. No.: 00021  
Linearity File No.: 00408 Lin. Date: 09/30/2014 Lin. No.: 00021  
Solution File No.: 00421 Soln. Date: 03/17/2015 Soln. No.: 00154  
Sequential File No.: 00421 File Date: 03/17/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
Control Solution %: 0.100% Expires: 08/22/2016  
Solution Control Lot: 14H132 Bottle No.: 0034

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:24S		
Control 1 EC	0.101%	15:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25S		
Control 2 EC	0.086%	15:26S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.097%	15:26S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:27S		
Control 3 EC	0.087%	15:28S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.095%	15:28S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:29S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 03/17/2015