

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00376	Calib. Date:	12/03/2010
Certification File No.:	00377	Cert. Date:	12/03/2010
Linearity File No.:	00378	Lin. Date:	12/03/2010
Solution File No.:	00405	Soln. Date:	03/29/2011
Sequential File No.:	00405	File Date:	03/29/2011

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUF S3-0060
Control Solution %:	0.100%			Expires:	12/16/2012
Solution Control Lot:	10L084			Bottle No.:	0140

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:51S		
Control 1 EC	-.---%	06:51S	20.6°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	06:51S	20.6°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 03/29/2011

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00376	Calib. No.: 00014
Certification File No.:	00377	Cert. No.: 00010
Linearity File No.:	00378	Lin. No.: 00010
Solution File No.:	00406	Soln. No.: 00097
Sequential File No.:	00406	File Date: 03/29/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0140

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:07S		
Control 1 EC	0.103%	17:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	17:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.008%	17:08S		
Control 2 EC	0.097%	17:09S	34.0°C	AMBIENT AIR CHECK ERR
Control 2 IR	-.---%	17:09S	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	17:10S		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 03/29/2011