

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00210 Calib. Date: 01/06/2009 Calib. No.: 00009
Certification File No.: 00211 Cert. Date: 01/06/2009 Cert. No.: 00007
Linearity File No.: 00212 Lin. Date: 01/06/2009 Lin. No.: 00007
Solution File No.: 00217 Soln. Date: 04/01/2009 Soln. No.: 00057
Sequential File No.: 00217 File Date: 04/01/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 01/09/2010
Solution Control Lot: 08A049 Bottle No.: 1008

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:36S		
Control 1 EC	0.100%	06:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:38S		
Control 2 EC	0.098%	06:38S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:39S		
Control 3 EC	0.098%	06:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 04/01/2009