

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00314 Calib. Date: 01/16/2012 Calib. No.: 00022
Certification File No.: 00315 Cert. Date: 01/16/2012 Cert. No.: 00014
Linearity File No.: 00316 Lin. Date: 01/16/2012 Lin. No.: 00014
Solution File No.: 00322 Soln. Date: 04/02/2012 Soln. No.: 00104
Sequential File No.: 00322 File Date: 04/02/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 09/15/2013
Solution Control Lot: 11I095 Bottle No.: 0444

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:00D		
Control 1 EC	0.101%	14:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	14:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		
Control 2 EC	0.100%	14:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:03D		
Control 3 EC	0.100%	14:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 04/02/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00314	Calib. Date: 01/16/2012
Certification File No.:	00315	Calib. No.: 00022
Linearity File No.:	00316	Cert. Date: 01/16/2012
Solution File No.:	00321	Cert. No.: 00014
Sequential File No.:	00321	Lin. Date: 01/16/2012
		Lin. No.: 00014
		Soln. Date: 04/02/2012
		Soln. No.: 00103
		File Date: 04/02/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	11I095	Expires: 09/15/2013
		Bottle No.: 0444

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:11D		
Control 1 EC	0.101%	11:11D	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	11:11D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	11:12D		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 04/02/2012

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00314	Calib. Date: 01/16/2012
Certification File No.:	00315	Calib. No.: 00022
Linearity File No.:	00316	Cert. Date: 01/16/2012
Solution File No.:	00320	Cert. No.: 00014
Sequential File No.:	00320	Lin. Date: 01/16/2012
		Lin. No.: 00014
		Soln. Date: 04/02/2012
		Soln. No.: 00102
		File Date: 04/02/2012

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUF S3-0060
Control Solution %: 0.100%		Expires: 09/15/2013
Solution Control Lot: 11I095		Bottle No.: 0444

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:26D		
Control 1 EC	-.---%	09:27D	22.5°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	09:27D	22.5°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOBSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 04/02/2012