

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00459	Calib. Date: 01/16/2012
Certification File No.:	00460	Calib. No.: 00019
Linearity File No.:	00461	Cert. Date: 01/16/2012
Solution File No.:	00478	Cert. No.: 00013
Sequential File No.:	00478	Lin. Date: 01/16/2012
		Lin. No.: 00013
		Soln. Date: 04/02/2012
		Soln. No.: 00113
		File Date: 04/02/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	110I95	Expires: 09/15/2013
		Bottle No.: 0439

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:52D		
Control 1 EC	0.099%	09:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53D		
Control 2 EC	0.099%	09:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:55D		
Control 3 EC	0.098%	09:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	09:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 04/02/2012