

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00244	Calib. Date: 12/21/2009
Certification File No.:	00245	Calib. No.: 00015
Linearity File No.:	00246	Cert. Date: 12/21/2009
Solution File No.:	00253	Cert. No.: 00009
Sequential File No.:	00253	Lin. Date: 12/21/2009
		Lin. No.: 00009
		Soln. Date: 04/08/2010
		Soln. No.: 00072
		File Date: 04/08/2010

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %: 0.100%		Expires: 12/01/2010
Solution Control Lot: 08L061		Bottle No.: 1056

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:35D		
Control 1 EC	0.101%	13:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36D		
Control 2 EC	0.100%	13:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:37D		
Control 3 EC	0.101%	13:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 04/08/2010