

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00346	Calib. No.: 00024
Certification File No.:	00347	Cert. No.: 00016
Linearity File No.:	00348	Lin. No.: 00016
Solution File No.:	00363	Soln. No.: 00123
Sequential File No.:	00363	File Date: 04/11/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0420

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:33D		
Control 1 EC	0.099%	08:34D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34D		
Control 2 EC	0.098%	08:35D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:36D		
Control 3 EC	0.098%	08:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_

Badge No.: 351

Date: 04/11/2013