

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00531	Calib. No.: 00025
Certification File No.:	00532	Cert. No.: 00015
Linearity File No.:	00533	Lin. No.: 00015
Solution File No.:	00575	Soln. No.: 00132
Sequential File No.:	00575	File Date: 04/11/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0419

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:22D		
Control 1 EC	0.100%	08:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:24D		
Control 2 EC	0.100%	08:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25D		
Control 3 EC	0.100%	08:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 04/11/2013