

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00443	Calib. Date: 02/17/2016
Certification File No.:	00444	Calib. No.: 00033
Linearity File No.:	00445	Cert. Date: 02/17/2016
Solution File No.:	00451	Cert. No.: 00024
Sequential File No.:	00451	Lin. Date: 02/17/2016
		Lin. No.: 00024
		Soln. Date: 04/13/2016
		Soln. No.: 00171
		File Date: 04/13/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 0085

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:09D		
Control 1 EC	0.099%	12:10D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:10D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		
Control 2 EC	0.098%	12:11D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:11D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12D		
Control 3 EC	0.098%	12:12D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:12D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  #421

Badge No.: 421

Date: 04/13/2016