

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00831	Calib. No.: 00032
Certification File No.:	00832	Cert. No.: 00022
Linearity File No.:	00833	Lin. No.: 00022
Solution File No.:	00843	Soln. No.: 00178
Sequential File No.:	00843	File Date: 04/13/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 0084

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:59D		
Control 1 EC	0.101%	12:00D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:00D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01D		
Control 2 EC	0.101%	12:01D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:01D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02D		
Control 3 EC	0.100%	12:03D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:03D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 04/13/2016