

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00287	Calib. Date: 03/22/2011
Certification File No.:	00288	Calib. No.: 00020
Linearity File No.:	00289	Cert. Date: 03/22/2011
Solution File No.:	00290	Cert. No.: 00012
Sequential File No.:	00290	Lin. Date: 03/22/2011
		Lin. No.: 00012
		Soln. Date: 03/22/2011
		Soln. No.: 00087
		File Date: 03/22/2011

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUM S3-0080
Control Solution %:	0.100%		Expires: 12/16/2012
Solution Control Lot:	10L084		Bottle No.: 0145

H (40) 4/14/11

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:16S	16.16 D (40)	
Control 1 EC	0.104%	15:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:17S		
Control 2 EC	0.103%	15:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	15:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18S		
Control 3 EC	0.103%	15:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	15:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19S	16.19 D (40)	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL # DDFP P2-0417 (40)

Changed By:

Last Name: DENNIS	First Name: MARC	MI: W.
Signature: <u>TPR. I [Signature] # 5925</u>	Badge No.: 5925	Date: 03/22/2011

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00287	Calib. Date: 03/22/2011
Certification File No.:	00288	Calib. No.: 00020
Linearity File No.:	00289	Cert. Date: 03/22/2011
Solution File No.:	00294	Cert. No.: 00012
Sequential File No.:	00294	Lin. Date: 03/22/2011
		Lin. No.: 00012
		Soln. Date: 04/14/2011
		Soln. No.: 00088
		File Date: 04/14/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0135

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:16D		
Control 1 EC	0.103%	10:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17D		
Control 2 EC	0.102%	10:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19D		
Control 3 EC	0.103%	10:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	10:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: _____

TAR. I [Signature] # 5925

Badge No.: 5925

Date: 04/14/2011