

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00020
Calibration File No.:	00287	Cert. Date:	03/22/2011
Certification File No.:	00288	Lin. Date:	03/22/2011
Linearity File No.:	00289	Soln. Date:	04/14/2011
Solution File No.:	00294	File Date:	04/14/2011
Sequential File No.:	00294		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	0135

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:16D		
Control 1 EC	0.103%	10:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17D		
Control 2 EC	0.102%	10:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19D		
Control 3 EC	0.103%	10:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	10:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: _____

T.P.R. I [Signature] # 5925

Badge No.: 5925

Date: 04/14/2011