

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00743	Calib. Date:	03/19/2015
Certification File No.:	00744	Cert. Date:	03/19/2015
Linearity File No.:	00745	Lin. Date:	03/19/2015
Solution File No.:	00752	Soln. Date:	04/16/2015
Sequential File No.:	00752	File Date:	04/16/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0060
Solution Control Lot:	14H132	Expires:	08/22/2016
		Bottle No.:	0038

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:35D		
Control 1 EC	0.101%	12:36D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:37D		
Control 2 EC	0.100%	12:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38D		
Control 3 EC	0.100%	12:39D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: 

Badge No.: 6040

Date: 04/16/2015