

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00235 Calib. Date: 01/06/2009 Calib. No.: 00010  
Certification File No.: 00236 Cert. Date: 01/06/2009 Cert. No.: 00006  
Linearity File No.: 00237 Lin. Date: 01/06/2009 Lin. No.: 00006  
Solution File No.: 00254 Soln. Date: 04/29/2009 Soln. No.: 00066  
Sequential File No.: 00254 File Date: 04/29/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 01/09/2010  
Solution Control Lot: 08A049 Bottle No.: 1010

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:18D		
Control 1 EC	0.100%	18:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:19D		
Control 2 EC	0.100%	18:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:21D		
Control 3 EC	0.100%	18:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: \_\_\_\_\_



Badge No.: 258

Date: 04/29/2009