

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00022
Calibration File No.:	00314	Calib. Date:	01/16/2012
Certification File No.:	00315	Cert. Date:	01/16/2012
Linearity File No.:	00316	Lin. Date:	01/16/2012
Solution File No.:	00324	Soln. Date:	04/30/2012
Sequential File No.:	00324	File Date:	04/30/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0060
Solution Control Lot:	11I095	Expires:	09/15/2013
		Bottle No.:	0434

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:27D		
Control 1 EC	0.100%	11:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		
Control 2 EC	0.099%	11:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30D		
Control 3 EC	0.100%	11:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 04/30/2012