

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00459	Calib. Date: 01/16/2012
Certification File No.:	00460	Calib. No.: 00019
Linearity File No.:	00461	Cert. Date: 01/16/2012
Solution File No.:	00480	Cert. No.: 00013
Sequential File No.:	00480	Lin. Date: 01/16/2012
		Lin. No.: 00013
		Soln. Date: 04/30/2012
		Soln. No.: 00114
		File Date: 04/30/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	11I095	Expires: 09/15/2013
		Bottle No.: 0429

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:22D		
Control 1 EC	0.098%	11:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 2 EC	0.099%	11:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		
Control 3 EC	0.099%	11:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature:  #351

Badge No.: 351

Date: 04/30/2012