

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00673	Calib. Date: 04/07/2014
Certification File No.:	00674	Cert. No.: 00018
Linearity File No.:	00675	Lin. Date: 04/07/2014
Solution File No.:	00679	Soln. No.: 00148
Sequential File No.:	00679	File Date: 05/01/2014

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1342

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:04D		
Control 1 EC	0.099%	13:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:06D		
Control 2 EC	0.098%	13:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:07D		
Control 3 EC	0.098%	13:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2 - 215 (MIG)

**Changed By:**

Last Name: GONCALVES      First Name: MICHELLE      MI: L

Signature: #1 Sgt. Helos #6040      Badge No.: 6040  
 Date: 05/01/2014