

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00244	Calib. Date: 12/21/2009
Certification File No.:	00245	Calib. No.: 00015
Linearity File No.:	00246	Cert. Date: 12/21/2009
Solution File No.:	00255	Cert. No.: 00009
Sequential File No.:	00255	Lin. Date: 12/21/2009
		Lin. No.: 00009
		Soln. Date: 05/07/2010
		Soln. No.: 00073
		File Date: 05/07/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	09I069	Expires: 09/18/2011
		Bottle No.: 0396

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:51D		
Control 1 EC	0.101%	12:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52D		
Control 2 EC	0.100%	12:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		
Control 3 EC	0.100%	12:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: _____



Badge No.: 258

Date: 05/07/2010