

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00307 Calib. Date: 12/21/2009 Calib. No.: 00012  
Certification File No.: 00308 Cert. Date: 12/21/2009 Cert. No.: 00008  
Linearity File No.: 00309 Lin. Date: 12/21/2009 Lin. No.: 00008  
Solution File No.: 00332 Soln. Date: 05/07/2010 Soln. No.: 00081  
Sequential File No.: 00332 File Date: 05/07/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 09/18/2011  
Solution Control Lot: 09I069 Bottle No.: 0391

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:47D		
Control 1 EC	0.101%	12:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:48D		
Control 2 EC	0.101%	12:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:49D		
Control 3 EC	0.101%	12:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: \_\_\_\_\_

*Michael Rogers*

Badge No.: 258

Date: 05/07/2010