

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00443	Calib. Date:	02/17/2016
Certification File No.:	00444	Cert. Date:	02/17/2016
Linearity File No.:	00445	Lin. Date:	02/17/2016
Solution File No.:	00452	Soln. Date:	05/12/2016
Sequential File No.:	00452	File Date:	05/12/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	08/04/2017
Solution Control Lot:	15H140			Bottle No.:	0089

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:14D		
Control 1 EC	0.099%	19:15D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:15D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:15D		
Control 2 EC	0.098%	19:16D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:16D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:17D		
Control 3 EC	0.098%	19:17D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:17D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 05/12/2016