

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00424 Calib. Date: 04/16/2015 Calib. No.: 00031  
Certification File No.: 00425 Cert. Date: 04/16/2015 Cert. No.: 00022  
Linearity File No.: 00426 Lin. Date: 04/16/2015 Lin. No.: 00022  
Solution File No.: 00428 Soln. Date: 05/13/2015 Soln. No.: 00157  
Sequential File No.: 00428 File Date: 05/13/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080  
Control Solution %: 0.100% Expires: 08/22/2016  
Solution Control Lot: 14H132 Bottle No.: 0029

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:06D		
Control 1 EC	0.101%	13:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:07D		
Control 2 EC	0.101%	13:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08D		
Control 3 EC	0.101%	13:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

*J. Blake #421*

Badge No.: 421

Date: 05/13/2015