

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00346	Calib. No.: 00024
Certification File No.:	00347	Cert. No.: 00016
Linearity File No.:	00348	Lin. No.: 00016
Solution File No.:	00364	Soln. No.: 00124
Sequential File No.:	00364	File Date: 05/14/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0417

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:37D		
Control 1 EC	0.099%	10:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:38D		
Control 2 EC	0.098%	10:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39D		
Control 3 EC	0.099%	10:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 05/14/2013