

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00577 Calib. Date: 04/17/2013 Calib. No.: 00026
Certification File No.: 00578 Cert. Date: 04/17/2013 Cert. No.: 00016
Linearity File No.: 00579 Lin. Date: 04/17/2013 Lin. No.: 00016
Solution File No.: 00588 Soln. Date: 05/14/2013 Soln. No.: 00134
Sequential File No.: 00588 File Date: 05/14/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 06/28/2014
Solution Control Lot: 12F102 Bottle No.: 0416

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:34D		
Control 1 EC	0.099%	10:35D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36D		
Control 2 EC	0.099%	10:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:37D		
Control 3 EC	0.099%	10:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:38D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 05/14/2013